



DIMENSIONS ARE IN INCHES AND ARE APPROXIMATE.

REVISIONS	BY	DATE	REVISIONS	BY	DATE
E	GAR	3-17-14	C	GAR	12-12-11
D	GAR	1-17-12	B	GAR	4-15-10
			A	DMR	6-30-08

E ADDED OPTIONAL 220-240 VAC INPUT POWER STEP-DOWN ISOLATION TRANSFORMER AND DISTRIBUTION JUNCTION BOX		GAR	3-17-14	C ADDED OPTIONAL MIXTURE SHUTOFF SOLENOID VALVE; MINOR GAS CHECK VALVE WAS IN UPFLOW POSITION NEXT TO MINOR GAS REGULATOR		GAR	12-12-11
D SURGE TANK SAFETY RELIEF VALVE WAS CONTROL DEVICES SF 50 ON OTHER END OF TANK; ADDED NOTES TO PIPE RELIEF VALVE AND TEST / VENT VALVE TO SAFE LOCATIONS; REMOVED WORD "GAS" (3 PLACES)		GAR	1-17-12	B UPDATED INLET FILTERS AND ADAPTERS TO CURRENT STANDARD; REPLACED ZERO / SAMPLE GAS SELECTOR VALVE WITH SWAGelok 5-WAY		GAR	4-15-10
				A HORN WAS OPTIONAL; REMOVED REFERENCES TO MODEL 7130 ANALOG ANALYZER		DMR	6-30-08

THERMCO		INSTRUMENT CORPORATION LA PORTE, INDIANA USA	
SCALE: N.T.S.	B. M. NO:	S. O. NO:	DRAWN BY: DB
DATE: 12-23-02	CUSTOMER:		APPROVED BY: GAR
TITLE: INSTRUCTION DRAWING FOR MODEL 6205 GAS MIXER WITH MODEL 9235 GAS ANALYZER			SUPERSEDES:
			DRAWING NUMBER: 1-12172